



Rupture of Membranes Test

INSTRUCTIONS FOR USE

ROM-5025, ROM-6025

ENGLISH INTENDED USE

The Clinical Innovations ROM Plus fetal membrane rupture test is a rapid, qualitative immunochromatographic test for the in vitro detection of amniotic fluid in vaginal secretions of pregnant women with signs and symptoms of ROM. The test detects AFP (alpha-fetoprotein) and IGFBP-1 (Insulin-like growth factor-binding protein-1 or PP12 (placental protein 12)) from amniotic fluid in vaginal secretions. The test is for prescription use by health care professionals to aid in the detection of rupture of membranes (ROM) in pregnant women in conjunction with other signs and symptoms.

REAGENTS AND COMPONENTS

Each ROM Plus kit contains:

- One ROM Plus test cassette with timer
- One sterile polyester vaginal swab
- One vial with buffer solution (phosphate-buffered saline)
- Directions for use (on each pouch)

The cassette test strip contains the following antibodies: sheep antihuman IGFBP-1, goat antihuman AFP, mouse conjugated colloid gold antihuman IGFBP-1 & AFP.

TEST SUMMARY

ROM Plus is a self-contained test kit that provides qualitative results for rupture of fetal membranes which is obtained at point-of-care sites and may be read in the lab. A speculum is not needed to obtain ROM Plus results. The test is non-invasive, with only a simple vaginal swab required. The sample is collected by placing the swab into the vagina for 15 seconds. The swab is then mixed into a vial containing buffer solution, and the diluted sample is applied to the sample pad of the test strip via the sample well on the cassette. A built-in timer is then activated and visualized as a convenient feature to indicate the time of the test. The liquid moves chromatographically and unidirectionally towards the absorbent pad.

TEST PRINCIPLE

During migration, the sample reacts with mono/polyclonal antibodies bound to the test strip membrane. These antibodies are immunoreactive to a combination of proteins, IGFBP-1 and AFP, which are markers of amniotic fluid. As the membrane absorbs the liquid sample, a control line will appear, indicating a sample was applied. If the sample contains the IGFBP-1 and/or AFP markers of amniotic fluid, it binds to the antibody of the test line, causing the test line to appear and indicating a positive result. If the sample does not contain the IGFBP-1 and/or AFP specific to amniotic fluid, only the control line will be visible, indicating a negative result.

STORAGE AND STABILITY

- Store the kit in a dry place at 4° to 37°C (40° to 99°F). Do not freeze.
- When stored in the foil pouch at the recommended temperature, the test is stable until the “Expiration” date on the pouch.
- Use ROM Plus within six (6) hours after opening foil pouch.
- Use ROM Plus within six (6) hours of collecting the vaginal swab sample and placing it into the buffer vial.

PRECAUTIONS, LIMITS AND WARNINGS

- ROM diagnoses should not be based on any single test.
- ROM Plus is for in vitro diagnostic use only.
- ROM Plus is for healthcare professional use only.
- Allow pouch containing ROM Plus to reach room temperature prior to utilizing the test.
- All instructions should be followed carefully for accurate results.
- Each ROM Plus test kit is single use and disposable and should not be reused.
- ROM Plus results are qualitative. No quantitative interpretations should be made.
- ROM Plus test kits will function properly with trace amounts of blood in the sample. Significant amounts of blood discharge may cause the test to malfunction and is not recommended.
- Safety precautions should be observed when collecting, handling, and disposing of test samples. Used test kits are biohazardous.
- Elevated fetal serum, urine, cord blood, and amniotic fluid as well as maternal serum levels of AFP have been reported in the literature in various developmental disorders such as neural-tube defects, hypothyroidism, autoimmune states, congenital heart defects, cystic fibrosis, etc. ROM Plus has not been evaluated for potential interference in these conditions.

Warning: The test may report positive results in patients with intact membranes (see specificity in the performance section) and therefore decisions to induce labor should not be based solely on the ROM Plus test results.

METHOD OR PERFORMANCE

- The ROM Plus assay was validated for the parameters of linearity, limit of detection, accuracy/reproducibility, sensitivity, specificity and cross-reactivity:
- High Concentration (“High Dose Hook” effect) – for the ROM Plus upper-detection range, the IGFBP-1 and AFP were tested. Concentrations of IGFBP-1 were tested up to 400,000 ng/ml and AFP up to 200,000 ng/ml with a positive visual results. (Although lines may appear lighter in presence of very high AFP and IGFBP-1 concentrations, any line is considered positive.)

- The lowest limit of detection (LOD) is 5 ng/ml for IGFBP-1, and 150 ng/ml for AFP. (These LOD concentrations refer to those in samples of vaginal secretions before dilution with buffer.)
- Reproducibility was tested on different days at six levels of amniotic fluid spiked into a negative control. The assay was run on three lots of ROM Plus to determine the visual positive or negative results. Two low positives, two moderate positives and two high positives were run on three lots of ROM Plus on four different days. No difference in activity was observed.
- To determine interference and cross-reactivity of the assay, Tylenol, aspirin, and three different bath products (KY Gel, Surgilube, Lever Soap, Noxema cream, Pert Shampoo), were spiked into the low positive control at a final concentration of 0.1% without visual loss of activity. The same products were spiked into the negative-matrix control and shown to be negative. In addition, human semen, urine and blood were spiked into the low positive at a 10% final concentration without loss of activity. Human semen, urine and blood were also spiked into the negative-control matrix and shown to be negative. IGFBP-1 assay does not cross-react with IGFBP-2, IGFBP-3, and IGFBP-4 based on Western Blot results. ROM Plus was shown to be negative when tested with specimens that were positive for bacterial vaginosis and sexually transmitted diseases. All samples were pH>4.5.

A multi-center prospective observational study was performed. Women included in the study were healthy pregnant women, 18 years of age or older, between 15-42 weeks of gestation reporting signs or symptoms of rupture of membranes (<34 ega [estimated gestational age], n=45; 34 to 24 ega, n=33; <24 ega, n=12). Patients with known placenta previa and active vaginal bleeding were excluded from the study. Initial evaluation included both the standard clinical assessment (Clin-Assess) for rupture of membranes comprising conventional clinical testing diagnosed by fluid leaking from the cervical os, or two of the following: pooling, positive nitrazine test, or ferning (SSE), as well as the new combination immunoassay ROM Plus containing the monoclonal and polyclonal antibodies to Insulin-like growth factor-binding protein-1 (IGFBP-1) and Alpha-fetoprotein (AFP).

Clinicians performing the SSE were blinded from the ROM Plus results. Clinicians performing ROM Plus testing were masked from Clin-Assess results. The clinical standard of pooling/ferning/nitrazine which has been shown in the literature to have sensitivities and specificities of 51-97% and 16-90% respectively (when these tests are used individually) are not an ideal reference standard. However, it was used as the reference standard in this study understanding its limitations and because it is commonly used in clinical hospital protocol to evaluate ROM.

NOTE: Corrections for ROM based on subsequent patient chart review are not reported here unlike other immunological studies evaluating for the detection of ruptured membranes with immunoassay technology.

Table 1: Clinical Study - ROM Plus vs. Clinical Assessment (pooling/ferning/nitrazine)

Combined	Clin-Assess	positive	negative	≥37	Clin-Assess	positive	negative
ROM Plus	positive	153	28	ROM Plus	positive	125	23
	negative	1	82		negative	1	33
Total: 264				Total: 182			
Sens	0.99	CI: 0.96 to 1.00		Sens	0.99	CI: 0.96 to 1.00	
Spec	0.75	CI: 0.66 to 0.82		Spec	0.58	CI: 0.46 to 0.71	
PPV	0.85	CI: 0.79 to 0.90		PPV	0.84	CI: 0.78 to 0.89	
NPV	0.99	CI: 0.94 to 1.00		NPV	0.97	CI: 0.85 to .099	
34-37 EGA	Clin-Assess	positive	negative	<34	Clin-Assess	positive	negative
ROM Plus	positive	18	3	ROM Plus	positive	10	2
	negative	0	16		negative	0	33
Total: 37				Total: 45			
Sens	1.00	CI: 0.82 to 1.00		Sens	1.00	CI: 0.72 to 1.00	
Spec	0.85	CI: 0.64 to 0.95		Spec	0.94	CI: 0.81 to 0.98	
PPV	0.87	CI: 0.65 to 0.95		PPV	0.83	CI: 0.55 to 0.95	
NPV	1.00	CI: 0.81 to 1.00		NPV	1.00	CI: 0.90 to 1.00	

Of the 264 patients in the study, rupture of membranes from Clin-Assess (SSE) as the standard compared to ROM Plus results were: sensitivity of 99%, specificity of 75%, positive predictive value of 85%, and negative predictive value of 99%. Table 1 shows 95% confidence levels for different gestational ages. Of the false positives 23 out of 28 (82%) were term and 5 of 28 (18%) were < 37 weeks gestation indicating the majority of the FP's were at term. Results in this study for pre-term patients were: sensitivity of 100%, specificity of 85%, positive predictive value of 87%, and negative predictive value of 100%. It is thought that false positives noted later in pregnancy and near delivery may be related to amniotic membranes beginning to leak (or becoming more permeable) without a gross rupture of membranes.

QUALITY CONTROL

Each ROM Plus test cassette contains Internal Quality Control, which verifies the integrity of the test procedure, verifies proper assembly of the test strip (device integrity) and that the test has been appropriately stored (not exposed to extreme temperature) (environment).

The positive control, demonstrated by the appearance of the control line (C), assures that the reagent is functioning properly (device integrity), an adequate sample volume was applied to the cassette (operator) and adequate capillary migration (lateral flow) occurred (device integrity). The negative control is represented by the lack of color along the entire strip length assuring no inappropriate binding occurred (device integrity). If the control line appears and there is color along the strip the test is functioning properly and does


ENGLISH not need to be repeated.

Because this test contains two levels (positive and negative) of internal quality control and is a qualitative, moderately complex test, it is the company's recommendation that external QC be completed for each new lot number or shipment of test material or if there is suspicion of improper storage and as required by regulatory or accrediting agencies.

ROM Plus Fetal Membrane Rupture Test Quality Control kit for external controls are available from Clinical Innovations, LLC. and are intended to monitor the performance of the product. The control test kits contain human IGFBP-1 and human AFP amniotic fluid proteins at concentrations of 20 ng/ml and 600 ng/ml. External negative controls that do not contain these proteins are also available in the Quality Control Kit. Always follow federal, state and local guidelines for quality control documentation.


Test Steps

- 1** Collect sample with vaginal swab




Remove ROM Plus contents from the packaging. Holding the buffer vial in an upright position, remove the shipping cap and set it aside. Remove the sterile swab from its package to collect a sample from the vagina. The tip of the swab should not touch anything prior to its insertion. Insert the swab tip into the vagina 2-3 inches (5-7 cm) deep. Withdraw the swab after a minimum of 15 seconds.


- 2** Mix and break the swab in buffer vial




Mix the swab in the buffer solution for 15 seconds.



Lift swab 1 cm from the bottom of the vial.



Bend the swab stick 90° in the forward direction and then 180° in the reverse direction.




Leave the tip in the vial and dispose of the remaining swab stick.

LAB - If using the test in the lab, place the shipping cap back on the buffer vial and send the vial to the lab. Allow at least 15 seconds for the swab to be in the buffer solution.

POC - If using the test POC, place the drop dispenser lid on the buffer vial, agitate the vial for 15 seconds.


3 Add sample and activate timer



Tear open the foil pouch and remove the ROM Plus cassette. Add 4-6 drops of the sample/ buffer solution to the sample well (Sample) of the cassette. Start the timer by firmly pressing and rolling thumb over the timer button from bottom to top. You will know the timer has been activated when you see the dark blue dye in the timer window.

A positive test result may be visible early (within 1-3 minutes) or may take the full 20 minutes. Please wait the full 20 minutes to confirm a negative result. Darkness of the lines may vary. The test is valid even if the lines are faint. Do not interpret test result based on darkness of the lines.

4 Read test result



If only a control line (C) is visible, the test result is negative. If both the control line (C) and test line (AF) are visible, the test result is positive. If no lines are visible, or just the test line (AF) is visible, the test result is invalid and should be repeated. A light visible line located in the test (AF) region should be considered positive; in addition, very high concentrations of proteins may result in a faint test (AF) line. It is recommended to read the strip by 20 minutes.

Bibliography

1. Thomasino T, Levi C. Diagnosing PROM with Combination Monoclonal/Polyclonal Immunologic Protein Detection. ACOG 2011 Annual Clinical Meeting, Washington DC, Poster Session, May 2011.
2. Caughey AB, Robinson JN, and Norwitz, ER. Contemporary diagnosis and management of preterm premature rupture of membranes. *Rev Obstet Gynecol.* 2008;1(1):11-22.
3. Tagore S, Kwek K. Comparative analysis of insulin-like growth factor binding protein-1 (IGFBP-1), placental alpha-microglobulin-1 (PAMG-1) and nitrazine test to diagnose premature rupture of membranes in pregnancy. *J Perinat Med* 2010;38:609-12.
4. El-Messidi A, Cameron A. Diagnosis of Premature Rupture of Membranes: Inspiration From the Past and Insights for the Future. *J.Obstet Gynaecol Can* 2010;32(6):561-69.
5. Albayrak, M, Ozdemir A, Koc O, Ankarali H, Ozen O. "Comparison of the diagnostic efficacy of the two rapid bedside immunoassays and combined clinical conventional diagnosis in prelabour rupture of membranes", *Eur.J. of Obstet. & Gyn. and Reprod. Biology.* 158 (2011)179-182.
6. Cousins LM, Smok DP, Lovett SM, et al. AmniSure placental alpha microglobulin-1 rapid immunoassay versus standard diagnostic methods for detection of rupture of membranes. *Am J Perinatol.* 2005;22(6):317-320.
7. Lee SE, Park JS, Norwitz ER, et al. Measurement of placental alpha-microglobulin-1 in cervicovaginal discharge to diagnose rupture of membranes. *Obstet Gynecol.* 2007;109(3):634-640.
8. Lee SM, Lee J, Seong HS, Lee SE, et al. The clinical significance of a false positive AmniSure test in women with term labor with intact membranes. *J. Matern Fetal Neonatal Med* 2009 Apr; 22(4): 305-310.
9. Chen FC, Dudenhausen JW. Comparison of two rapid strip tests based on IGFBP-1 and PAMG-1 for the detection of amniotic fluid. *Am J Perinatol.* 2008;25(4):243-246.
10. Yamada H, Kishida T, Negishi H, Sagawa T, Yamaguchi M, Sato C, Nakamura I, Sato H, Sakai K, Yamaguchi T, Fujimoto S. Comparison of an improved AFP kit with the intra-amniotic PSP dye-injection method in equivocal cases of preterm premature rupture of the fetal membranes. *J Obstet Gynaecol Res.* 1997 Jun;23(3):307-11.



Clinical Innovations
747 West 4170 South
Murray, Utah USA 84123
P. 801-268-8200
Toll Free: 888-268-6222

clinicalinnovations.com

Australia Sponsor:
Clinical Innovations Australia
36 Matheson Street,
Virginia, Qld, Australia
Phone: +(07) 3862 2000



**CLINICAL
INNOVATIONS**

forMOM. forBABY. forLIFE.

© Copyright 2020, Clinical Innovations, LLC.
All rights reserved. ART-0087 Rev. 02